

## PRODUCT DATA SHEET

### Integral IN 230 T

The new generation of Integral T process thermostats for professional external temperature control from -30 to 120 °C: The new LAUDA Integral T process thermostats ensure efficient control of external temperature control processes in the temperature range from -30 to 150 °C. The units enable rapid temperature changes due to adapted heating and cooling capacities with small internal volumes. Due to the open hydraulic system, the unit vents quickly and without functional restrictions. This makes it ideal for temperature control processes with frequent consumer or specimen changes. Classic areas of application are reaction controls or climate simulations. A reliable, powerful submersible pump and the internal bypass for pressure limitation are standard features of the Integral T.



Working temperature min.  
-30 °C



Working temperature max.  
120 °C

#### Technical Attributes

Working temperature min.	
Working temperature max.	
Ambient temperature min.	
Ambient temperature max.	
Temperature stability	
Filling volume min.	
Filling volume max.	
Heater power	
Cooling output at 20°C measured with ethanol	
Cooling output at 10°C measured with ethanol	
Cooling output at 0°C measured with ethanol	
Cooling output at -10°C measured with ethanol	
Cooling output at -20°C measured with ethanol	
Cooling output at -30°C measured with ethanol	
Power consumption	

#### Integral IN 230 T

-30 °C
120 °C
5 °C
40 °C
0.05 ±K
3.6 L
8.7 L
2.5 kW
2.2 kW
1.9 kW
1.5 kW
1.0 kW
0.6 kW
0.15 kW
3.7 kW

LAUDA DR. R. WOBSER GMBH & CO. KG  
Pfarrstraße 41/43 · 97922 Lauda-Königshofen  
Postfach 1251 · 97912 Lauda-Königshofen · DE

T +49 (0) 9343 503-0 · F +49 (0) 9343 503-222  
info@lauda.de · www.lauda.de  
WEEE-Reg-Nr.: DE 66 42 40 57

Kommanditgesellschaft: Sitz Lauda-Königshofen  
Registergericht Mannheim · HRA 560069

Persönlich haftende Gesellschafterin:  
LAUDA DR. R. WOBSER Verwaltungs-GmbH  
Sitz Lauda-Königshofen  
Registergericht Mannheim · HRB 560226

Geschäftsführer:  
Dr. Gunther Wobser (Vors.), Dr. Mario Englert  
Dr. Marc Stricker  
Beirat: Dr. Gerhard Wobser



## PRODUCT DATA SHEET

### Integral IN 230 T

#### Technical Attributes

Tube size	<u>3/4"</u>
Pump connection thread	<u>G 3/4</u>
Pump pressure max.	<u>3.5 bar</u>
Pump flow max. (pressure)	<u>40 L/min</u>
Dimensions (WxDxH) in mm	<u>430x550x760</u>

#### Integral IN 230 T

<u>3/4"</u>
<u>G 3/4</u>
<u>3.5 bar</u>
<u>40 L/min</u>
<u>430x550x760</u>

LAUDA DR. R. WOBSER GMBH & CO. KG  
Pfarrstraße 41/43 · 97922 Lauda-Königshofen  
Postfach 1251 · 97912 Lauda-Königshofen · DE

T +49 (0) 9343 503-0 · F +49 (0) 9343 503-222  
[info@lauda.de](mailto:info@lauda.de) · [www.lauda.de](http://www.lauda.de)  
WEEE-Reg-Nr.: DE 66 42 40 57

Kommanditgesellschaft: Sitz Lauda-Königshofen  
Registergericht Mannheim · HRA 560069

Persönlich haftende Gesellschafterin:  
LAUDA DR. R. WOBSER Verwaltungs-GmbH  
Sitz Lauda-Königshofen  
Registergericht Mannheim · HRB 560226

Geschäftsführer:  
Dr. Gunther Wobser (Vors.), Dr. Mario Englert  
Dr. Marc Stricker  
Beirat: Dr. Gerhard Wobser